


STENT Graft


2021.5.1～2021.10.31




	n	max	median	75%Tile
2室EVAR	13	2434	1156	1918
2室TEVAR	4	2170	825	1796
9室EVAR	20	7067	1533	1952
9室TEVAR	1	7123	—	—

TAVI


2021.5.1～2021.10.31



	n	max	median	75%Tile
TAVI TF	35	2476	862	986
TAVI TA	3	715	442	520




手技	n	Pka(Gycm ²)			Ka.r(mGy)		
		max	median	75%Tile	max	median	75%Tile
診断カテ	90	181	81	108	857	416	556
血管内治療	58	306	155	186	3117	1040	1545




循環器内科

2021.5.1～2021.10.31




	n	max	50%Tile	75%Tile
1室CAG	151	1949	636	872
3室CAG	260	2026	465	639
1室PCI	38	5591	911	1269
3室PCI	163	4834	884	1532
1室PTA	18	1597	226	343
3室PTA	2	119	101	110
Ablation	110	343	79	129
Ablation 3.75	29	158	67	96




	n	max	median	75%Tile
C-TACE	16	1179	521	721
TAI	2	159	153	—
TAE	4	837	593	752
B-RTO	3	146	550	802
BAE	4	816	708	


装置基準透視線量（循内）



装置基準透視線量（胸腹部）



装置基準透視線量（脳外）



装置基準透視線量（StentGraft・PTA）

